



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/18/2024

Lab File ID: BM044654.D Time Analyzed: 11:50

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	181490	7.698	729361	10.48	422995	14.34
UPPER LIMIT	362980	8.198	1458720	10.975	845990	14.839
LOWER LIMIT	90745	7.198	364681	9.975	211498	13.839
EPA SAMPLE NO.						
01 MW-01RE	202767	7.70	772524	10.48	428984	14.34

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.