



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/15/2024

Lab File ID: BF137610.D Time Analyzed: 15:34

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	518790	11.245	315670	13.874	268004	15.298
UPPER LIMIT	1037580	11.745	631340	14.374	536008	15.798
LOWER LIMIT	259395	10.745	157835	13.374	134002	14.798
EPA SAMPLE NO.						
01 PB159613BL	572257	11.24	350600	13.87	288694	15.30
02 PB159613BS	592768	11.25	363148	13.88	315730	15.30
03 MW-01	476309	11.24	380092	13.87	326161	15.30
04 PB159613BSD	614466	11.25	368786	13.88	320365	15.30

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.