



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/18/2024

Lab File ID: BM044654.D Time Analyzed: 11:50

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	831298	17.098	768575	21.297	882881	23.544
UPPER LIMIT	1662600	17.598	1537150	21.797	1765760	24.044
LOWER LIMIT	415649	16.598	384288	20.797	441441	23.044
EPA SAMPLE NO.						
01 MW-01RE	830911	17.10	796584	21.30	943523	23.54

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.