



Surrogate Summary

SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P1747-01	MW-01	2-Fluorophenol	150	63.7	42		10	139
		Phenol-d6	150	38.1	25		10	134
		Nitrobenzene-d5	100	93.5	93		49	133
		2-Fluorobiphenyl	100	90.6	91		52	132
		2,4,6-Tribromophenol	150	149	99		32	145
		Terphenyl-d14	100	92.3	92		36	145
P1747-02	MW-01-DUP	2-Fluorophenol	150	60.5	40		10	139
		Phenol-d6	150	35.7	24		10	134
		Nitrobenzene-d5	100	92.7	93		49	133
		2-Fluorobiphenyl	100	90.1	90		52	132
		2,4,6-Tribromophenol	150	137	92		32	145
		Terphenyl-d14	100	86.1	86		36	145
P1747-04	MW-02	2-Fluorophenol	150	49.2	33		10	139
		Phenol-d6	150	28.4	19		10	134
		Nitrobenzene-d5	100	90.1	90		49	133
		2-Fluorobiphenyl	100	85.2	85		52	132
		2,4,6-Tribromophenol	150	103	69		32	145
		Terphenyl-d14	100	90.3	90		36	145
P1747-05	MW-04	2-Fluorophenol	150	60.7	40		10	139
		Phenol-d6	150	35.4	24		10	134
		Nitrobenzene-d5	100	94.5	94		49	133
		2-Fluorobiphenyl	100	89.7	90		52	132
		2,4,6-Tribromophenol	150	126	84		32	145
		Terphenyl-d14	100	88.3	88		36	145
PB159586BL	PB159586BL	2-Fluorophenol	150	147	98		10	139
		Phenol-d6	150	138	92		10	134
		Nitrobenzene-d5	100	86.8	87		49	133
		2-Fluorobiphenyl	100	84.9	85		52	132
		2,4,6-Tribromophenol	150	135	90		32	145
		Terphenyl-d14	100	86.7	87		36	145
PB159586BS	PB159586BS	2-Fluorophenol	150	148	99		10	139
		Phenol-d6	150	139	93		10	134
		Nitrobenzene-d5	100	85.3	85		49	133
		2-Fluorobiphenyl	100	86.9	87		52	132
		2,4,6-Tribromophenol	150	141	94		32	145
		Terphenyl-d14	100	82.6	83		36	145
PB159586BSD	PB159586BSD	2-Fluorophenol	150	144	96		10	139
		Phenol-d6	150	136	91		10	134
		Nitrobenzene-d5	100	84.3	84		49	133
		2-Fluorobiphenyl	100	84.1	84		52	132
		2,4,6-Tribromophenol	150	139	92		32	145
		Terphenyl-d14	100	82.2	82		36	145