



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/14/2024

Lab File ID: BG060632.D Time Analyzed: 10:25

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	100214	8.322	455197	11.17	300034	14.95
UPPER LIMIT	200428	8.822	910394	11.666	600068	15.45
LOWER LIMIT	50107	7.822	227599	10.666	150017	14.45
EPA SAMPLE NO.						
01 MW-01	87894	8.32	407330	11.17	278240	14.95
02 MW-01-DUP	95427	8.32	445954	11.17	304535	14.95
03 MW-02	92610	8.32	433473	11.17	306273	14.94
04 MW-04	92415	8.33	426428	11.17	303077	14.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.