



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/14/2024

Lab File ID: BG060649.D Time Analyzed: 22:39

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	96079	8.323	434457	11.17	283666	14.95
UPPER LIMIT	192158	8.823	868914	11.667	567332	15.451
LOWER LIMIT	48039.5	7.823	217229	10.667	141833	14.451
EPA SAMPLE NO.						
01 PB159586BL	87880	8.32	398324	11.17	273634	14.95
02 PB159586BS	109222	8.33	513453	11.17	325000	14.95
03 PB159586BSD	104909	8.32	497758	11.17	316713	14.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.