



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG No.: P1747
Instrument ID: MSVOA_X Calibration Date/Time: 03/14/2024 09:43
Lab File ID: VX040630.D Init. Calib. Date(s): 03/05/2024 03/05/2024
Heated Purge: (Y/N) N Init. Calib. Time(s): 10:58 12:30
GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Methyl tert-Butyl Ether	6.602	6.116		-7.36	
Carbon Tetrachloride	0.551	0.479	0.1	-13.07	
Chloroform	4.158	3.929	0.2	-5.51	
1,1,1-Trichloroethane	0.654	0.584	0.1	-10.7	
Benzene	1.402	1.224	0.5	-12.7	
Toluene	1.652	1.432	0.4	-13.32	
Tetrachloroethene	0.356	0.336	0.2	-5.62	
Ethyl Benzene	1.905	1.799	0.1	-5.56	
m/p-Xylenes	0.704	0.671	0.3	-4.69	
o-Xylene	0.677	0.620	0.3	-8.42	
1,4-Dichlorobenzene	0.862	0.825	0.2	-4.29	
1,2-Dichloroethane-d4	2.996	3.235	0.01	7.98	
Toluene-d8	1.326	1.232	0.01	-7.09	
4-Bromofluorobenzene	0.515	0.520	0.2	0.97	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.