



Surrogate Summary

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P1747-03	MW-01	1,2-Dichloroethane-d4	30	30.6	102	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	29.1	97	63	112
VX0314WBL01	VX0314WBL01	1,2-Dichloroethane-d4	30	30.4	101	91	110
		Toluene-d8	30	27.7	92	91	112
		4-Bromofluorobenzene	30	26.1	87	63	112
VX0314WBS01	VX0314WBS01	1,2-Dichloroethane-d4	30	29.8	99	91	110
		Toluene-d8	30	30.8	103	91	112
		4-Bromofluorobenzene	30	31.2	104	63	112
VX0314WBSD0	VX0314WBSD01	1,2-Dichloroethane-d4	30	27.6	92	91	110
		Toluene-d8	30	33.1	110	91	112
		4-Bromofluorobenzene	30	30.7	102	63	112