



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: SW624.1

Datafile : VX040631.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VX0314WBS01	Methyl tert-butyl Ether	20	18.8	ug/L	94			82	114	
	Carbon Tetrachloride	20	19.0	ug/L	95			70	130	
	Chloroform	20	19.1	ug/L	96			70	135	
	1,1,1-Trichloroethane	20	18.8	ug/L	94			70	130	
	Benzene	20	18.8	ug/L	94			65	135	
	Toluene	20	19.9	ug/L	100			70	130	
	Tetrachloroethene	20	20.2	ug/L	101			70	130	
	Ethyl Benzene	20	20.1	ug/L	101			60	140	
	m/p-Xylenes	40	40.4	ug/L	101			87	111	
	o-Xylene	20	20.3	ug/L	102			87	111	
	1,4-Dichlorobenzene	20	20.4	ug/L	102			65	135	