



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: SW624.1

Datafile : VX040632.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VX0314WBSD01	Methyl tert-butyl Ether	20	19.4	ug/L	97	3		82	114	20
	Carbon Tetrachloride	20	20.2	ug/L	101	6		70	130	20
	Chloroform	20	19.7	ug/L	99	3		70	135	20
	1,1,1-Trichloroethane	20	19.9	ug/L	100	6		70	130	20
	Benzene	20	18.3	ug/L	92	2		65	135	20
	Toluene	20	21.2	ug/L	106	6		70	130	20
	Tetrachloroethene	20	21.6	ug/L	108	7		70	130	20
	Ethyl Benzene	20	19.4	ug/L	97	4		60	140	20
	m/p-Xylenes	40	37.8	ug/L	95	6		87	111	20
	o-Xylene	20	19.6	ug/L	98	4		87	111	20
	1,4-Dichlorobenzene	20	21.0	ug/L	105	3		65	135	20