



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG No.: P1747
Instrument ID: MSVOA_N Calibration Date/Time: 03/14/2024 11:22
Lab File ID: VN081401.D Init. Calib. Date(s): 03/05/2024 03/05/2024
Heated Purge: (Y/N) N Init. Calib. Time(s): 12:00 13:59
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	48.66	2.676003	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.