



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN030524**

|                          |               |                                                       |                      |  |              |  |  |
|--------------------------|---------------|-------------------------------------------------------|----------------------|--|--------------|--|--|
| Review By                | John Carlone  | Review On                                             | 3/6/2024 9:39:37 AM  |  |              |  |  |
| Supervise By             | Mahesh Dadoda | Supervise On                                          | 3/6/2024 11:26:59 AM |  |              |  |  |
| SubDirectory             | VN030524      | HP Acquire Method                                     | HP Processing Method |  | 82N030524W.M |  |  |
| STD. NAME                |               | STD REF.#                                             |                      |  |              |  |  |
| Tune/Reschk              |               | VP126377                                              |                      |  |              |  |  |
| Initial Calibration Stds |               | VP126395,VP126396,VP126397,VP126398,VP126399,VP126400 |                      |  |              |  |  |
| CCC                      |               | VP126378,VP126379                                     |                      |  |              |  |  |
| Internal Standard/PEM    |               |                                                       |                      |  |              |  |  |
| ICV/I.BLK                |               | VP126401                                              |                      |  |              |  |  |
| Surrogate Standard       |               |                                                       |                      |  |              |  |  |
| MS/MSD Standard          |               |                                                       |                      |  |              |  |  |
| LCS Standard             |               |                                                       |                      |  |              |  |  |

| Sr# | SampleId    | ClientID    | Data File Name | Date-Time         | Comment      | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|--------------|----------|--------|
| 1   | BFB         | BFB         | VN081302.D     | 05 Mar 2024 09:31 |              | JC\MD    | Ok     |
| 2   | VSTDCCC050  | VSTDCCC050  | VN081303.D     | 05 Mar 2024 11:20 | Need ICAL    | JC\MD    | Not Ok |
| 3   | VSTDICC100  | VSTDICC100  | VN081304.D     | 05 Mar 2024 12:00 | Good for DOD | JC\MD    | Ok,M   |
| 4   | VSTDICCC050 | VSTDICCC050 | VN081305.D     | 05 Mar 2024 12:24 | LR- 20       | JC\MD    | Ok,M   |
| 5   | VSTDICC020  | VSTDICC020  | VN081306.D     | 05 Mar 2024 12:48 |              | JC\MD    | Ok,M   |
| 6   | VSTDICC010  | VSTDICC010  | VN081307.D     | 05 Mar 2024 13:12 |              | JC\MD    | Ok,M   |
| 7   | VSTDICC005  | VSTDICC005  | VN081308.D     | 05 Mar 2024 13:36 |              | JC\MD    | Ok,M   |
| 8   | VSTDICC001  | VSTDICC001  | VN081309.D     | 05 Mar 2024 13:59 |              | JC\MD    | Ok,M   |
| 9   | IBLK        | IBLK        | VN081310.D     | 05 Mar 2024 14:48 |              | JC\MD    | Ok     |
| 10  | VSTDICV050  | ICVVN030524 | VN081311.D     | 05 Mar 2024 15:15 |              | JC\MD    | Ok,M   |

M : Manual Integration



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_N

## Daily Analysis Runlog For Sequence/QC Batch ID # VN031424

| Review By                                                                                                                                         | John Carlone  | Review On            | 3/15/2024 9:22:48 AM |                   |               |          |        |
|---------------------------------------------------------------------------------------------------------------------------------------------------|---------------|----------------------|----------------------|-------------------|---------------|----------|--------|
| Supervise By                                                                                                                                      | Mahesh Dadoda | Supervise On         | 3/15/2024 4:04:49 PM |                   |               |          |        |
| SubDirectory                                                                                                                                      | VN031424      | HP Acquire Method    | HP Processing Method |                   | 82N030524W.M  |          |        |
| STD. NAME                                                                                                                                         |               | STD REF.#            |                      |                   |               |          |        |
| Tune/Reschk<br>Initial Calibration Stds<br><br>CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard |               | VP126585             |                      |                   |               |          |        |
|                                                                                                                                                   |               | VP126586,VP126587    |                      |                   |               |          |        |
|                                                                                                                                                   |               |                      |                      |                   |               |          |        |
|                                                                                                                                                   |               |                      |                      |                   |               |          |        |
|                                                                                                                                                   |               |                      |                      |                   |               |          |        |
| Sr#                                                                                                                                               | SampleId      | ClientID             | Data File Name       | Date-Time         | Comment       | Operator | Status |
| 1                                                                                                                                                 | BFB           | BFB                  | VN081400.D           | 14 Mar 2024 10:49 |               | JC\MD    | Ok     |
| 2                                                                                                                                                 | VSTDCCC050    | VSTDCCC050           | VN081401.D           | 14 Mar 2024 11:22 | pH#Lot#V12668 | JC\MD    | Ok,M   |
| 3                                                                                                                                                 | VN0314MBL01   | VN0314MBL01          | VN081402.D           | 14 Mar 2024 11:56 |               | JC\MD    | Ok     |
| 4                                                                                                                                                 | VN0314WBL01   | VN0314WBL01          | VN081403.D           | 14 Mar 2024 12:20 |               | JC\MD    | Ok     |
| 5                                                                                                                                                 | VN0314WBS01   | VN0314WBS01          | VN081404.D           | 14 Mar 2024 12:44 |               | JC\MD    | Ok,M   |
| 6                                                                                                                                                 | VN0314WBSD01  | VN0314WBSD01         | VN081405.D           | 14 Mar 2024 13:18 |               | JC\MD    | Ok,M   |
| 7                                                                                                                                                 | P1725-02      | RT-3860RE            | VN081406.D           | 14 Mar 2024 13:41 | vial B pH#5.0 | JC\MD    | Ok     |
| 8                                                                                                                                                 | P1727-01      | 240313044-02-VOA     | VN081407.D           | 14 Mar 2024 14:05 | vial B pH#6.0 | JC\MD    | Ok,M   |
| 9                                                                                                                                                 | IBLK          | IBLK                 | VN081408.D           | 14 Mar 2024 14:29 |               | JC\MD    | Ok     |
| 10                                                                                                                                                | P1727-02      | 240313039-03-TRIP-BL | VN081409.D           | 14 Mar 2024 14:53 | vial B pH#6.0 | JC\MD    | Ok     |
| 11                                                                                                                                                | P1747-06      | TRIP-BLANK-1         | VN081410.D           | 14 Mar 2024 15:17 | vial A pH<2   | JC\MD    | Ok     |
| 12                                                                                                                                                | P1749-05      | TRIP-BLANK           | VN081411.D           | 14 Mar 2024 15:41 | vial A pH<2   | JC\MD    | Ok     |
| 13                                                                                                                                                | P1747-01      | MW-01                | VN081412.D           | 14 Mar 2024 16:05 | vial A pH<2   | JC\MD    | Ok     |
| 14                                                                                                                                                | P1747-02      | MW-01-DUP            | VN081413.D           | 14 Mar 2024 16:29 | vial A pH<2   | JC\MD    | Ok     |
| 15                                                                                                                                                | P1747-04      | MW-02                | VN081414.D           | 14 Mar 2024 16:53 | vial A pH<2   | JC\MD    | Ok     |
| 16                                                                                                                                                | P1747-05      | TWP-04               | VN081415.D           | 14 Mar 2024 17:17 | vial A pH<2   | JC\MD    | Ok     |
| 17                                                                                                                                                | P1749-01      | MW-1                 | VN081416.D           | 14 Mar 2024 17:41 | vial A pH<2   | JC\MD    | Ok     |
| 18                                                                                                                                                | P1749-02      | MW-2A                | VN081417.D           | 14 Mar 2024 18:05 | vial A pH<2   | JC\MD    | Ok     |
| 19                                                                                                                                                | P1749-03      | MW-2                 | VN081418.D           | 14 Mar 2024 18:29 | vial A pH<2   | JC\MD    | Ok     |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN031424**

|                                                                                                    |               |                   |                      |                   |                                       |       |          |  |
|----------------------------------------------------------------------------------------------------|---------------|-------------------|----------------------|-------------------|---------------------------------------|-------|----------|--|
| Review By                                                                                          | John Carlone  | Review On         | 3/15/2024 9:22:48 AM |                   |                                       |       |          |  |
| Supervise By                                                                                       | Mahesh Dadoda | Supervise On      | 3/15/2024 4:04:49 PM |                   |                                       |       |          |  |
| SubDirectory                                                                                       | VN031424      | HP Acquire Method | HP Processing Method |                   | 82N030524W.M                          |       |          |  |
| STD. NAME                                                                                          |               | STD REF.#         |                      |                   |                                       |       |          |  |
| Tune/Reschk<br>Initial Calibration Stds                                                            |               | VP126585          |                      |                   |                                       |       |          |  |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard |               | VP126586,VP126587 |                      |                   |                                       |       |          |  |
| 20                                                                                                 | P1749-04      | MW-4              | VN081419.D           | 14 Mar 2024 18:53 | vial A pH<2                           | JC\MD | Ok,M     |  |
| 21                                                                                                 | P1769-01      | MW-2              | VN081420.D           | 14 Mar 2024 19:17 | vial A pH<2 Need 50X                  | JC\MD | Dilution |  |
| 22                                                                                                 | P1769-02      | MW-1              | VN081421.D           | 14 Mar 2024 19:41 | vial A pH<2 E flag in previous sample | JC\MD | ReRun    |  |
| 23                                                                                                 | P1769-03      | FB-3-14-24        | VN081422.D           | 14 Mar 2024 20:05 | vial A pH<2 carryover conc., FB       | JC\MD | Not Ok   |  |
| 24                                                                                                 | P1769-07      | TB-3-14-24        | VN081423.D           | 14 Mar 2024 20:29 | vial A pH<2 carryover conc., TB       | JC\MD | Not Ok   |  |
| 25                                                                                                 | VSTDCCC050    | VSTDCCC050EC      | VN081424.D           | 14 Mar 2024 20:52 |                                       | JC\MD | Ok,M     |  |

M : Manual Integration