



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: SW8260-Low

Datafile : VN081404.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VN0314WBS01	Dichlorodifluoromethane	20	20.2	ug/L	101			69	116	
	Chloromethane	20	18.4	ug/L	92			65	116	
	Vinyl chloride	20	18.1	ug/L	91			65	117	
	Bromomethane	20	18.5	ug/L	93			58	125	
	Chloroethane	20	19.2	ug/L	96			56	128	
	Trichlorofluoromethane	20	19.6	ug/L	98			73	115	
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99			80	112	
	1,1-Dichloroethene	20	19.1	ug/L	96			74	110	
	Acetone	100	100	ug/L	100			60	125	
	Carbon disulfide	20	17.0	ug/L	85			64	112	
	Methyl tert-butyl Ether	20	19.1	ug/L	96			78	114	
	Methyl Acetate	20	23.5	ug/L	117			67	125	
	Methylene Chloride	20	20.4	ug/L	102			72	114	
	trans-1,2-Dichloroethene	20	17.6	ug/L	88			75	108	
	1,1-Dichloroethane	20	20.0	ug/L	100			78	112	
	Cyclohexane	20	18.9	ug/L	95			75	110	
	2-Butanone	100	100	ug/L	100			65	122	
	Carbon Tetrachloride	20	20.4	ug/L	102			77	113	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			77	110	
	Bromochloromethane	20	22.0	ug/L	110			70	124	
	Chloroform	20	20.8	ug/L	104			79	113	
	1,1,1-Trichloroethane	20	20.6	ug/L	103			80	108	
	Methylcyclohexane	20	18.3	ug/L	92			72	115	
	Benzene	20	19.2	ug/L	96			82	109	
	1,2-Dichloroethane	20	20.6	ug/L	103			80	115	
	Trichloroethene	20	19.0	ug/L	95			77	113	
	1,2-Dichloropropane	20	19.7	ug/L	99			83	111	
	Bromodichloromethane	20	20.5	ug/L	103			83	110	
	4-Methyl-2-Pentanone	100	100	ug/L	100			74	118	
	Toluene	20	19.8	ug/L	99			82	110	
	t-1,3-Dichloropropene	20	19.7	ug/L	99			79	110	
	cis-1,3-Dichloropropene	20	19.7	ug/L	99			82	110	
	1,1,2-Trichloroethane	20	20.0	ug/L	100			83	112	
	2-Hexanone	100	100	ug/L	100			73	117	
	Dibromochloromethane	20	21.0	ug/L	105			82	110	
	1,2-Dibromoethane	20	19.8	ug/L	99			81	110	
	Tetrachloroethene	20	19.7	ug/L	99			67	123	
	Chlorobenzene	20	19.8	ug/L	99			82	109	
	Ethyl Benzene	20	19.3	ug/L	97			83	109	
	m/p-Xylenes	40	38.2	ug/L	96			82	110	
	o-Xylene	20	18.9	ug/L	95			83	109	
	Styrene	20	20.1	ug/L	101			80	111	
	Bromoform	20	20.0	ug/L	100			79	109	
	Isopropylbenzene	20	19.8	ug/L	99			83	112	
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98			76	118	
	1,3-Dichlorobenzene	20	18.6	ug/L	93			82	108	
	1,4-Dichlorobenzene	20	19.1	ug/L	96			82	107	
	1,2-Dichlorobenzene	20	19.0	ug/L	95			82	109	
	1,2-Dibromo-3-Chloropropane	20	18.8	ug/L	94			68	112	



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: P1747
Client: LiRo Engineers, Inc.
Analytical Method: SW8260-Low

Datafile : VN081404.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VN0314WBS01	1,2,4-Trichlorobenzene	20	18.7	ug/L	94			55	133	
	1,2,3-Trichlorobenzene	20	18.6	ug/L	93			59	129	