



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747
Lab File ID: VN081401.D Date Analyzed: 03/14/2024
Instrument ID: MSVOA_N Time Analyzed: 11:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	340353	8.22	595569	9.10	547809	11.87
UPPER LIMIT	680706	8.724	1191140	9.6	1095620	12.365
LOWER LIMIT	170177	7.724	297785	8.6	273905	11.365
EPA SAMPLE NO.						
MW-01	303621	8.23	551652	9.11	483758	11.87
MW-01-DUP	297408	8.23	548071	9.11	467622	11.87
MW-02	287608	8.23	536366	9.11	459785	11.87
TWP-04	293736	8.23	538371	9.11	468090	11.87
TRIP-BLANK-1	344974	8.22	628116	9.11	532846	11.87
VN0314WBL01	317492	8.23	583725	9.11	501625	11.87
VN0314WBS01	299990	8.23	553441	9.11	502583	11.87
VN0314WBSD01	294169	8.23	534919	9.11	486415	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.