



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747
Lab File ID: VN081401.D Date Analyzed: 03/14/2024
Instrument ID: MSVOA_N Time Analyzed: 11:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	252603	13.794				
UPPER LIMIT	505206	14.294				
LOWER LIMIT	126302	13.294				
EPA SAMPLE NO.						
MW-01	192317	13.79				
MW-01-DUP	179160	13.79				
MW-02	179683	13.79				
TWP-04	182546	13.79				
TRIP-BLANK-1	205915	13.79				
VN0314WBL01	187883	13.79				
VN0314WBS01	222348	13.79				
VN0314WBSD01	221293	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.