

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG No.: P2403

Calibration Sequence : FE050924		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	127759977	127760	FE048049.D
500	64227450	128455	FE048050.D
200	25352809	126764	FE048051.D
100	14338926	143389	FE048052.D
50	7178823	143576	FE048053.D
AVG RF : 133989		% RSD : 6.484	AVG RT : 15.0528