

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG No.: P2403

Calibration Sequence : FF041124		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	123957688	123958	FF014100.D
500	59859134	119718	FF014101.D
200	23438012	117190	FF014102.D
100	11810325	118103	FF014103.D
50	5402626	108053	FF014104.D
AVG RF : 117404		% RSD : 4.972	AVG RT : 14.9444