

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG No.: P2403

Calibration Sequence : FG050924		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	123374033	123374	FG013761.D
500	61273408	122547	FG013762.D
200	25869928	129350	FG013763.D
100	13985859	139859	FG013764.D
50	7135604	142712	FG013765.D
AVG RF : 131568		% RSD : 7.073	AVG RT : 15.0686