

Continuing Calibration Report for SequenceID : FC050924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC066356.D

Aliphatic C9-C12	8065172.000	60.000	3.221	6.507	134419.533	152832.209	12.048
Aliphatic C12-C16	5508672.000	40.000	6.508	9.904	137716.800	152122.872	9.470
Aliphatic C16-C21	8182633.000	60.000	9.905	13.269	136377.217	145403.822	6.208
Aliphatic C21-C28	10394973.000	80.000	13.270	16.930	129937.163	137872.665	5.756
Aliphatic C28-C40	14201632.000	120.000	16.931	21.821	118346.933	118607.818	0.220
Aliphatic EPH	46353082.000	360.000	3.221	21.821	128758.561	136782.856	5.866