

Continuing Calibration Report for SequenceID : FC051024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC066387.D

Aliphatic C9-C12	7590576.000	60.000	3.213	6.503	126509.600	152832.209	17.223
Aliphatic C12-C16	5132981.000	40.000	6.504	9.902	128324.525	152122.872	15.644
Aliphatic C16-C21	7555550.000	60.000	9.903	13.267	125925.833	145403.822	13.396
Aliphatic C21-C28	9503645.000	80.000	13.268	16.929	118795.563	137872.665	13.837
Aliphatic C28-C40	14405724.000	120.000	16.930	21.822	120047.700	118607.818	-1.214
Aliphatic EPH	44188476.000	360.000	3.213	21.822	122745.767	136782.856	10.262