

Continuing Calibration Report for SequenceID : FD050924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047682.D

Aromatic C10-C12	7130907.000	40.000	4.175	5.899	178272.675	165442.694	-7.755
Aromatic C12-C16	10640646.000	60.000	5.900	8.512	177344.100	167174.051	-6.084
Aromatic C16-C21	12806244.000	80.000	8.513	12.782	160078.050	160639.062	0.349
Aromatic C21-C36	26994204.000	180.000	12.783	18.189	149967.800	147764.997	-1.491
Aromatic EPH	57572001.000	360.000	4.175	18.189	159922.225	155824.931	-2.629