

Continuing Calibration Report for SequenceID : FD050924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047692.D

Aromatic C10-C12	7587529.000	40.000	4.175	5.899	189688.225	165442.694	-14.655
Aromatic C12-C16	11330566.000	60.000	5.900	8.512	188842.767	167174.051	-12.962
Aromatic C16-C21	13627484.000	80.000	8.513	12.782	170343.550	160639.062	-6.041
Aromatic C21-C36	28639590.000	180.000	12.783	18.189	159108.833	147764.997	-7.677
Aromatic EPH	61185169.000	360.000	4.175	18.189	169958.803	155824.931	-9.070