

Continuing Calibration Report for SequenceID : FD051024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047694.D

Aromatic C10-C12	7677819.000	40.000	4.174	5.899	191945.475	165442.694	-16.019
Aromatic C12-C16	11470980.000	60.000	5.900	8.513	191183.000	167174.051	-14.362
Aromatic C16-C21	13952579.000	80.000	8.514	12.783	174407.238	160639.062	-8.571
Aromatic C21-C36	28585459.000	180.000	12.784	18.191	158808.106	147764.997	-7.473
Aromatic EPH	61686837.000	360.000	4.174	18.191	171352.325	155824.931	-9.965