

Continuing Calibration Report for SequenceID : FD051024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047701.D

Aromatic C10-C12	7276830.000	40.000	4.174	5.899	181920.750	165442.694	-9.960
Aromatic C12-C16	10866732.000	60.000	5.900	8.513	181112.200	167174.051	-8.338
Aromatic C16-C21	13349406.000	80.000	8.514	12.783	166867.575	160639.062	-3.877
Aromatic C21-C36	27398255.000	180.000	12.784	18.191	152212.528	147764.997	-3.010
Aromatic EPH	58891223.000	360.000	4.174	18.191	163586.731	155824.931	-4.981