

Continuing Calibration Report for SequenceID : FD051024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047713.D

Aromatic C10-C12	7400959.000	40.000	4.174	5.899	185023.975	165442.694	-11.836
Aromatic C12-C16	11025946.000	60.000	5.900	8.513	183765.767	167174.051	-9.925
Aromatic C16-C21	13376931.000	80.000	8.514	12.783	167211.638	160639.062	-4.092
Aromatic C21-C36	27297904.000	180.000	12.784	18.191	151655.022	147764.997	-2.633
Aromatic EPH	59101740.000	360.000	4.174	18.191	164171.500	155824.931	-5.356