

Continuing Calibration Report for SequenceID : FD051024AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047727.D

Aromatic C10-C12	7444145.000	40.000	4.174	5.899	186103.625	165442.694	-12.488
Aromatic C12-C16	11068004.000	60.000	5.900	8.513	184466.733	167174.051	-10.344
Aromatic C16-C21	13394644.000	80.000	8.514	12.783	167433.050	160639.062	-4.229
Aromatic C21-C36	27592389.000	180.000	12.784	18.191	153291.050	147764.997	-3.740
Aromatic EPH	59499182.000	360.000	4.174	18.191	165275.506	155824.931	-6.065