

Continuing Calibration Report for SequenceID : FD051324AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047737.D

Aliphatic C9-C12	12015870.000	60.000	3.087	6.340	200264.500	196375.066	-1.981
Aliphatic C12-C16	8127862.000	40.000	6.341	9.716	203196.550	198002.288	-2.623
Aliphatic C16-C21	12130706.000	60.000	9.717	13.061	202178.433	196888.857	-2.687
Aliphatic C21-C28	15095137.000	80.000	13.062	16.702	188689.213	183433.041	-2.865
Aliphatic C28-C40	19488760.000	120.000	16.703	21.437	162406.333	157961.790	-2.814
Aliphatic EPH	66858335.000	360.000	3.087	21.437	185717.597	180961.070	-2.628