

Continuing Calibration Report for SequenceID : FE051024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE048102.D

Aliphatic C9-C12	8794761.000	60.000	3.249	6.862	146579.350	131281.972	-11.652
Aliphatic C12-C16	5825951.000	40.000	6.863	10.298	145648.775	135027.408	-7.866
Aliphatic C16-C21	8517671.000	60.000	10.299	13.661	141961.183	135438.387	-4.816
Aliphatic C21-C28	10654894.000	80.000	13.662	17.319	133186.175	133066.004	-0.090
Aliphatic C28-C40	15912939.000	120.000	17.320	22.223	132607.825	137525.037	3.576
Aliphatic EPH	49706216.000	360.000	3.249	22.223	138072.822	134868.341	-2.376