

Continuing Calibration Report for SequenceID : FE051024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE048112.D

Aliphatic C9-C12	8842617.000	60.000	3.249	6.862	147376.950	131281.972	-12.260
Aliphatic C12-C16	5816553.000	40.000	6.863	10.298	145413.825	135027.408	-7.692
Aliphatic C16-C21	8464612.000	60.000	10.299	13.661	141076.867	135438.387	-4.163
Aliphatic C21-C28	10695031.000	80.000	13.662	17.319	133687.888	133066.004	-0.467
Aliphatic C28-C40	16287502.000	120.000	17.320	22.223	135729.183	137525.037	1.306
Aliphatic EPH	50106315.000	360.000	3.249	22.223	139184.208	134868.341	-3.200