

Continuing Calibration Report for SequenceID : FF051224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014219.D

Aliphatic C9-C12	8429950.000	60.000	3.140	6.752	140499.167	131877.277	-6.538
Aliphatic C12-C16	5635607.000	40.000	6.753	10.185	140890.175	130041.750	-8.342
Aliphatic C16-C21	8689691.000	60.000	10.186	13.550	144828.183	128542.899	-12.669
Aliphatic C21-C28	10815903.000	80.000	13.551	17.213	135198.788	118156.858	-14.423
Aliphatic C28-C40	11898579.000	120.000	17.214	22.066	99154.825	102475.050	3.240
Aliphatic EPH	45469730.000	360.000	3.140	22.066	126304.806	118267.875	-6.796