

Continuing Calibration Report for SequenceID : FG051024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG013807.D

Aliphatic C9-C12	8109873.000	60.000	3.226	6.850	135164.550	135383.675	0.162
Aliphatic C12-C16	5612270.000	40.000	6.851	10.296	140306.750	139843.467	-0.331
Aliphatic C16-C21	8612341.000	60.000	10.297	13.673	143539.017	141841.298	-1.197
Aliphatic C21-C28	11001534.000	80.000	13.674	17.345	137519.175	135646.148	-1.381
Aliphatic C28-C40	14205271.000	120.000	17.346	22.294	118377.258	118727.001	0.295
Aliphatic EPH	47541289.000	360.000	3.226	22.294	132059.136	131461.581	-0.455