

Continuing Calibration Report for SequenceID : FG051024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG013817.D

Aliphatic C9-C12	8374666.000	60.000	3.226	6.850	139577.767	135383.675	-3.098
Aliphatic C12-C16	5706609.000	40.000	6.851	10.296	142665.225	139843.467	-2.018
Aliphatic C16-C21	8532147.000	60.000	10.297	13.673	142202.450	141841.298	-0.255
Aliphatic C21-C28	10666423.000	80.000	13.674	17.345	133330.288	135646.148	1.707
Aliphatic C28-C40	13927720.000	120.000	17.346	22.294	116064.333	118727.001	2.243
Aliphatic EPH	47207565.000	360.000	3.226	22.294	131132.125	131461.581	0.251