

Continuing Calibration Report for SequenceID : FG051024AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG013823.D

Aliphatic C9-C12	7731081.000	60.000	3.226	6.850	128851.350	135383.675	4.825
Aliphatic C12-C16	5220844.000	40.000	6.851	10.296	130521.100	139843.467	6.666
Aliphatic C16-C21	7722965.000	60.000	10.297	13.673	128716.083	141841.298	9.253
Aliphatic C21-C28	9651682.000	80.000	13.674	17.345	120646.025	135646.148	11.058
Aliphatic C28-C40	13055677.000	120.000	17.346	22.294	108797.308	118727.001	8.363
Aliphatic EPH	43382249.000	360.000	3.226	22.294	120506.247	131461.581	8.333