

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P2403 **SAS No.:** P2403 **SDG NO.:** P2403

Continuing Calib Date: 05/08/2024 **Initial Calibration Date(s):** 05/06/2024 05/06/2024

Continuing Calib Time: 21:30 **Initial Calibration Time(s):** 11:33 13:10

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.46	7.47	7.37	7.57	0.01
2,4-DCAA	7.27	7.28	7.18	7.38	0.01
DICHLORPROP	8.17	8.17	8.07	8.27	0.00
2,4-D	8.40	8.40	8.30	8.50	0.00
2,4,5-TP(Silvex)	9.28	9.28	9.18	9.38	0.01
2,4,5-T	9.57	9.57	9.47	9.67	0.00
2,4-DB	10.14	10.15	10.05	10.25	0.01
Dinoseb	11.36	11.36	11.26	11.46	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/08/2024 Initial Calibration Date(s): 05/06/2024 05/06/2024

Continuing Calib Time: 21:30 Initial Calibration Time(s): 11:33 13:10

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.98	7.99	7.89	8.09	0.01
2,4-DCAA	7.78	7.78	7.68	7.88	0.00
DICHLORPROP	8.70	8.70	8.60	8.80	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.94	9.94	9.84	10.04	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.92	10.93	10.83	11.03	0.01
Dinoseb	11.30	11.30	11.20	11.40	0.00