

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL01 Date Analyzed: 05/08/2024

Lab Sample No.: HSTDCCC750 Data File : PS026673.D Time Analyzed: 10:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.568	9.474	9.674	712.540	712.500	0.0
2,4,5-TP(Silvex)	9.277	9.181	9.381	706.060	712.500	-0.9
2,4-D	8.399	8.303	8.503	714.320	705.000	1.3
2,4-DB	10.143	10.048	10.248	718.850	712.500	0.9
2,4-DCAA	7.275	7.177	7.377	710.390	750.000	-5.3
DICAMBA	7.463	7.365	7.565	701.300	705.000	-0.5
DICHLORPROP	8.168	8.072	8.272	689.250	705.000	-2.2
Dinoseb	11.355	11.262	11.462	707.030	705.000	0.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL01 Date Analyzed: 05/08/2024

Lab Sample No.: HSTDCCC750 Data File : PS026673.D Time Analyzed: 10:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.355	10.259	10.459	721.820	712.500	1.3
2,4,5-TP(Silvex)	9.935	9.838	10.038	724.070	712.500	1.6
2,4-D	9.030	8.932	9.132	706.150	705.000	0.2
2,4-DB	10.921	10.825	11.025	709.810	712.500	-0.4
2,4-DCAA	7.782	7.684	7.884	746.490	750.000	-0.5
DICAMBA	7.982	7.885	8.085	716.310	705.000	1.6
DICHLORPROP	8.698	8.602	8.802	704.000	705.000	-0.1
Dinoseb	11.301	11.204	11.404	725.110	705.000	2.9