

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/08/2024

Lab Sample No.: HSTDCCC750 Data File : PS026684.D Time Analyzed: 21:30

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.567	9.474	9.674	720.800	712.500	1.2
2,4,5-TP(Silvex)	9.275	9.181	9.381	712.690	712.500	0.0
2,4-D	8.398	8.303	8.503	709.040	705.000	0.6
2,4-DB	10.142	10.048	10.248	723.550	712.500	1.6
2,4-DCAA	7.274	7.177	7.377	744.180	750.000	-0.8
DICAMBA	7.461	7.365	7.565	716.060	705.000	1.6
DICHLORPROP	8.167	8.072	8.272	701.940	705.000	-0.4
Dinoseb	11.355	11.262	11.462	721.720	705.000	2.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/08/2024

Lab Sample No.: HSTDCCC750 Data File : PS026684.D Time Analyzed: 21:30

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.356	10.259	10.459	758.080	712.500	6.4
2,4,5-TP(Silvex)	9.935	9.838	10.038	748.380	712.500	5.0
2,4-D	9.029	8.932	9.132	728.820	705.000	3.4
2,4-DB	10.921	10.825	11.025	744.190	712.500	4.4
2,4-DCAA	7.781	7.684	7.884	764.010	750.000	1.9
DICAMBA	7.982	7.885	8.085	733.000	705.000	4.0
DICHLORPROP	8.698	8.602	8.802	724.280	705.000	2.7
Dinoseb	11.300	11.204	11.404	766.930	705.000	8.8