

Manual Integration Report

Sequence:	PS050624	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDCCC750	PS026661.D	DCPA	Abdul	5/7/2024 8:20:22 AM	Ankita	5/7/2024 11:32:33	Peak Integrated by Software incorrectly
HSTDCCC750	PS026661.D	Picloram	Abdul	5/7/2024 8:20:22 AM	Ankita	5/7/2024 11:32:33	Peak Integrated by Software incorrectly
HSTDCCC750	PS026670.D	DCPA	Abdul	5/7/2024 8:20:46 AM	Ankita	5/7/2024 11:32:42	Peak Integrated by Software incorrectly
HSTDCCC750	PS026670.D	Picloram	Abdul	5/7/2024 8:20:46 AM	Ankita	5/7/2024 11:32:42	Peak Integrated by Software incorrectly

Manual Integration Report

Sequence:	PS050824	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDCCC750	PS026673.D	2,4-DCAA	Abdul	5/9/2024 9:47:54 AM	Ankita	5/9/2024 1:58:55	Peak Integrated by Software incorrectly
HSTDCCC750	PS026673.D	DINOSEB	Abdul	5/9/2024 9:47:54 AM	Ankita	5/9/2024 1:58:55	Peak Integrated by Software incorrectly
HSTDCCC750	PS026673.D	Pentachlorophenol	Abdul	5/9/2024 9:47:54 AM	Ankita	5/9/2024 1:58:55	Peak Integrated by Software incorrectly
PB160776BL	PS026674.D	2,4-DCAA	Abdul	5/9/2024 9:47:57 AM	Ankita	5/9/2024 1:58:56	Peak Integrated by Software incorrectly
PB160776BL	PS026674.D	2,4-DCAA #2	Abdul	5/9/2024 9:47:57 AM	Ankita	5/9/2024 1:58:56	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	2,4,5-T #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	2,4,5-TP (SILVEX) #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	2,4-DB	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	2,4-DB #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	2,4-DCAA #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	4-Nitrophenol #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	Dalapon	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	Dalapon #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly

Manual Integration Report

Sequence:	PS050824	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
P2403-01	PS026676.D	DCPA	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	DICAMBA #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	DINOSEB	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	MCPA #2	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	MCPD	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026676.D	Picloram	Abdul	5/9/2024 3:29:33 PM	Ankita	5/9/2024 4:08:31	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	2,4,5-TP (SILVEX) #2	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	2,4-DB	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	2,4-DB #2	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	2,4-DCAA #2	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	4-Nitrophenol	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	4-Nitrophenol #2	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	Dalapon	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly

Manual Integration Report

Sequence:

PS050824

Instrument

ECD_s

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
P2403-01	PS026677.D	Dalapon #2	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	DINOSEB	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	MCPP	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
P2403-01	PS026677.D	Picloram	Abdul	5/9/2024 9:48:09 AM	Ankita	5/9/2024 1:58:59	Peak Integrated by Software incorrectly
HSTDCCC750	PS026684.D	Picloram	Abdul	5/9/2024 9:48:35 AM	Ankita	5/9/2024 2:01:44	Peak Integrated by Software incorrectly
HSTDCCC750	PS026696.D	DINOSEB #2	Abdul	5/9/2024 3:29:40 PM	Ankita	5/9/2024 4:08:35	Peak Integrated by Software incorrectly
HSTDCCC750	PS026696.D	Picloram	Abdul	5/9/2024 3:29:40 PM	Ankita	5/9/2024 4:08:35	Peak Integrated by Software incorrectly