



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

LOD-MDL-WATER-01-Q

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Lab Sample ID: P2403-07 Date(s) Analyzed: 05/09/2024 05/09/2024

Instrument ID (1): ECD_R Instrument ID (2): ECD_R

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PR066589.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.872	4.822	4.922	0.26	0.27	6.22
COLUMN 1	2	4.894	4.844	4.944	0.25		
	3	4.959	4.909	5.009	0.32		
	4	5.06	5.01	5.11	0.27		
	5	5.374	5.324	5.424	0.26		
	1	4.005	3.955	4.055	0.24		
COLUMN 2	2	4.022	3.972	4.072	0.29	0.26	
	3	4.2	4.15	4.25	0.22		
	4	4.25	4.2	4.3	0.25		
	5	4.466	4.416	4.516	0.28		
Aroclor-1260	1	6.584	6.534	6.634	0.30	0.29	17.25
COLUMN 1	2	6.865	6.815	6.915	0.28		
	3	7.254	7.204	7.304	0.30		
	4	7.496	7.446	7.546	0.30		
	5	7.834	7.784	7.884	0.25		
	1	5.552	5.502	5.602	0.26		
COLUMN 2	2	5.756	5.706	5.806	0.24	0.24	
	3	5.914	5.864	5.964	0.25		
	4	6.417	6.367	6.467	0.23		
	5	6.682	6.632	6.732	0.21		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

LOD-MDL-WATER-01-Q

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Lab Sample ID: P2403-07 Date(s) Analyzed: 05/09/2024 05/09/2024

Instrument ID (1): ECD_R Instrument ID (2): ECD_R

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PR066590.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1221	1	3.874	3.824	3.924	0.30	0.34	13.17
COLUMN 1	2	3.963	3.913	4.013	0.49		
	3	4.041	3.991	4.091	0.22		
	4	0	0	0	0		
	5	0	0	0	0		
	1	3.118	3.068	3.168	0.20		
COLUMN 2	2	3.196	3.146	3.246	0.43	0.30	
	3	3.281	3.231	3.331	0.26		
	4	0	0	0	0		
	5	0	0	0	0		