

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/15/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 16:48 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.62	5.62	5.52	5.72	0.00
Aroclor-1016-2	(2)	5.64	5.64	5.54	5.74	0.00
Aroclor-1016-3	(3)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-4	(4)	5.80	5.80	5.70	5.90	0.00
Aroclor-1016-5	(5)	6.10	6.10	6.00	6.20	0.00
Aroclor-1260-1	(1)	7.22	7.22	7.12	7.32	0.00
Aroclor-1260-2	(2)	7.48	7.47	7.37	7.57	-0.01
Aroclor-1260-3	(3)	7.84	7.83	7.73	7.93	-0.01
Aroclor-1260-4	(4)	8.06	8.06	7.96	8.16	0.00
Aroclor-1260-5	(5)	8.38	8.38	8.28	8.48	0.00
Tetrachloro-m-xylene		4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl		10.21	10.20	10.10	10.30	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/15/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 16:48 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.66	4.67	4.57	4.77	0.01
Aroclor-1016-2	(2)	4.68	4.69	4.59	4.79	0.01
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.90	4.91	4.81	5.01	0.02
Aroclor-1016-5	(5)	5.11	5.12	5.02	5.22	0.01
Aroclor-1260-1	(1)	6.13	6.14	6.04	6.24	0.01
Aroclor-1260-2	(2)	6.32	6.33	6.23	6.43	0.01
Aroclor-1260-3	(3)	6.47	6.48	6.38	6.58	0.01
Aroclor-1260-4	(4)	6.94	6.95	6.85	7.05	0.01
Aroclor-1260-5	(5)	7.18	7.19	7.09	7.29	0.01
Tetrachloro-m-xylene		3.58	3.59	3.49	3.69	0.01
Decachlorobiphenyl		8.54	8.56	8.46	8.66	0.02