

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/17/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 15:27 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.47	6.47	6.37	6.57	0.00
Aroclor-1254-2 (2)	6.69	6.69	6.59	6.79	0.00
Aroclor-1254-3 (3)	7.06	7.05	6.95	7.15	0.00
Aroclor-1254-4 (4)	7.34	7.34	7.24	7.44	0.00
Aroclor-1254-5 (5)	7.76	7.75	7.65	7.85	-0.01
Tetrachloro-m-xylene	4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl	10.21	10.20	10.10	10.30	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/17/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 15:27 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.45	5.46	5.36	5.56	0.01
Aroclor-1254-2 (2)	5.60	5.61	5.51	5.71	0.01
Aroclor-1254-3 (3)	6.00	6.01	5.91	6.11	0.01
Aroclor-1254-4 (4)	6.22	6.23	6.13	6.33	0.01
Aroclor-1254-5 (5)	6.64	6.65	6.55	6.75	0.01
Tetrachloro-m-xylene	3.58	3.59	3.49	3.69	0.01
Decachlorobiphenyl	8.54	8.56	8.46	8.66	0.02