

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/17/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 19:34 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.62       | 5.62      | 5.52      | 5.72  | 0.00       |
| Aroclor-1016-2       | (2) | 5.64       | 5.64      | 5.54      | 5.74  | 0.00       |
| Aroclor-1016-3       | (3) | 5.70       | 5.70      | 5.60      | 5.80  | 0.00       |
| Aroclor-1016-4       | (4) | 5.80       | 5.80      | 5.70      | 5.90  | 0.00       |
| Aroclor-1016-5       | (5) | 6.10       | 6.10      | 6.00      | 6.20  | 0.01       |
| Aroclor-1260-1       | (1) | 7.22       | 7.22      | 7.12      | 7.32  | 0.00       |
| Aroclor-1260-2       | (2) | 7.47       | 7.47      | 7.37      | 7.57  | 0.00       |
| Aroclor-1260-3       | (3) | 7.83       | 7.83      | 7.73      | 7.93  | 0.00       |
| Aroclor-1260-4       | (4) | 8.06       | 8.06      | 7.96      | 8.16  | 0.00       |
| Aroclor-1260-5       | (5) | 8.38       | 8.38      | 8.28      | 8.48  | 0.00       |
| Tetrachloro-m-xylene |     | 4.45       | 4.45      | 4.35      | 4.55  | 0.00       |
| Decachlorobiphenyl   |     | 10.21      | 10.20     | 10.10     | 10.30 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/17/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 19:34 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.66       | 4.67      | 4.57      | 4.77 | 0.01       |
| Aroclor-1016-2       | (2) | 4.68       | 4.69      | 4.59      | 4.79 | 0.01       |
| Aroclor-1016-3       | (3) | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1016-4       | (4) | 4.89       | 4.91      | 4.81      | 5.01 | 0.02       |
| Aroclor-1016-5       | (5) | 5.10       | 5.12      | 5.02      | 5.22 | 0.02       |
| Aroclor-1260-1       | (1) | 6.13       | 6.14      | 6.04      | 6.24 | 0.01       |
| Aroclor-1260-2       | (2) | 6.31       | 6.33      | 6.23      | 6.43 | 0.02       |
| Aroclor-1260-3       | (3) | 6.47       | 6.48      | 6.38      | 6.58 | 0.01       |
| Aroclor-1260-4       | (4) | 6.93       | 6.95      | 6.85      | 7.05 | 0.02       |
| Aroclor-1260-5       | (5) | 7.17       | 7.19      | 7.09      | 7.29 | 0.02       |
| Tetrachloro-m-xylene |     | 3.58       | 3.59      | 3.49      | 3.69 | 0.01       |
| Decachlorobiphenyl   |     | 8.54       | 8.56      | 8.46      | 8.66 | 0.02       |