

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/17/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 19:51 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.62       | 5.62      | 5.52      | 5.72  | 0.00       |
| Aroclor-1242-2 (2)   | 5.64       | 5.64      | 5.54      | 5.74  | 0.00       |
| Aroclor-1242-3 (3)   | 5.70       | 5.70      | 5.60      | 5.80  | 0.00       |
| Aroclor-1242-4 (4)   | 5.80       | 5.80      | 5.70      | 5.90  | 0.00       |
| Aroclor-1242-5 (5)   | 6.54       | 6.54      | 6.44      | 6.64  | 0.01       |
| Tetrachloro-m-xylene | 4.45       | 4.45      | 4.35      | 4.55  | 0.00       |
| Decachlorobiphenyl   | 10.21      | 10.20     | 10.10     | 10.30 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/17/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 19:51 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.66       | 4.67      | 4.57      | 4.77 | 0.01       |
| Aroclor-1242-2 (2)   | 4.68       | 4.69      | 4.59      | 4.79 | 0.01       |
| Aroclor-1242-3 (3)   | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1242-4 (4)   | 4.93       | 4.95      | 4.85      | 5.05 | 0.02       |
| Aroclor-1242-5 (5)   | 5.45       | 5.46      | 5.36      | 5.56 | 0.01       |
| Tetrachloro-m-xylene | 3.58       | 3.59      | 3.49      | 3.69 | 0.01       |
| Decachlorobiphenyl   | 8.54       | 8.56      | 8.46      | 8.66 | 0.02       |