

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/18/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 00:50 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.62	5.62	5.52	5.72	0.00
Aroclor-1242-2 (2)	5.64	5.64	5.54	5.74	0.00
Aroclor-1242-3 (3)	5.70	5.70	5.60	5.80	0.00
Aroclor-1242-4 (4)	5.80	5.80	5.70	5.90	0.00
Aroclor-1242-5 (5)	6.54	6.54	6.44	6.64	0.01
Tetrachloro-m-xylene	4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl	10.21	10.20	10.10	10.30	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/18/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 00:50 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.66	4.67	4.57	4.77	0.01
Aroclor-1242-2 (2)	4.68	4.69	4.59	4.79	0.01
Aroclor-1242-3 (3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1242-4 (4)	4.93	4.95	4.85	5.05	0.02
Aroclor-1242-5 (5)	5.45	5.46	5.36	5.56	0.01
Tetrachloro-m-xylene	3.58	3.59	3.49	3.69	0.01
Decachlorobiphenyl	8.54	8.56	8.46	8.66	0.02