

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/18/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 05:48 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.62	5.62	5.52	5.72	0.00
Aroclor-1248-2 (2)	5.89	5.89	5.79	5.99	0.00
Aroclor-1248-3 (3)	6.10	6.09	5.99	6.19	0.00
Aroclor-1248-4 (4)	6.50	6.50	6.40	6.60	0.00
Aroclor-1248-5 (5)	6.54	6.54	6.44	6.64	0.01
Tetrachloro-m-xylene	4.45	4.45	4.35	4.55	0.00
Decachlorobiphenyl	10.20	10.20	10.10	10.30	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/18/2024 Initial Calibration Date(s): 05/06/2024 05/07/2024

Continuing Calib Time: 05:48 Initial Calibration Time(s): 16:26 00:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.66	4.67	4.57	4.77	0.01
Aroclor-1248-2 (2)	4.89	4.90	4.80	5.00	0.01
Aroclor-1248-3 (3)	4.93	4.95	4.85	5.05	0.02
Aroclor-1248-4 (4)	5.10	5.12	5.02	5.22	0.02
Aroclor-1248-5 (5)	5.49	5.50	5.40	5.60	0.01
Tetrachloro-m-xylene	3.58	3.59	3.49	3.69	0.01
Decachlorobiphenyl	8.54	8.56	8.46	8.66	0.02