

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/14/2024 Initial Calibration Date(s): 05/13/2024 05/13/2024

Continuing Calib Time: 21:34 Initial Calibration Time(s): 16:05 23:26

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.42	5.43	5.33	5.53	0.01
Aroclor-1016-2	(2)	5.44	5.45	5.35	5.55	0.01
Aroclor-1016-3	(3)	5.50	5.51	5.41	5.61	0.01
Aroclor-1016-4	(4)	5.60	5.61	5.51	5.71	0.01
Aroclor-1016-5	(5)	5.90	5.91	5.81	6.01	0.01
Aroclor-1260-1	(1)	7.03	7.04	6.94	7.14	0.01
Aroclor-1260-2	(2)	7.29	7.29	7.19	7.39	0.00
Aroclor-1260-3	(3)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-4	(4)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-5	(5)	8.18	8.19	8.09	8.29	0.01
Tetrachloro-m-xylene		4.25	4.26	4.16	4.36	0.01
Decachlorobiphenyl		9.91	9.91	9.81	10.01	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/14/2024 Initial Calibration Date(s): 05/13/2024 05/13/2024

Continuing Calib Time: 21:34 Initial Calibration Time(s): 16:05 23:26

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.43	4.43	4.33	4.53	0.00
Aroclor-1016-2	(2)	4.44	4.45	4.35	4.55	0.01
Aroclor-1016-3	(3)	4.61	4.62	4.52	4.72	0.01
Aroclor-1016-4	(4)	4.66	4.67	4.57	4.77	0.01
Aroclor-1016-5	(5)	4.86	4.87	4.77	4.97	0.01
Aroclor-1260-1	(1)	5.88	5.89	5.79	5.99	0.01
Aroclor-1260-2	(2)	6.07	6.08	5.98	6.18	0.01
Aroclor-1260-3	(3)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-4	(4)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-5	(5)	6.93	6.94	6.84	7.04	0.01
Tetrachloro-m-xylene		3.37	3.37	3.27	3.47	0.00
Decachlorobiphenyl		8.26	8.27	8.17	8.37	0.01