

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/10/2024 Initial Calibration Date(s): 04/23/2024 04/23/2024

Continuing Calib Time: 02:05 Initial Calibration Time(s): 10:25 17:00

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.55       | 4.55      | 4.45      | 4.65 | 0.00       |
| Aroclor-1016-2       | (2) | 4.57       | 4.57      | 4.47      | 4.67 | 0.00       |
| Aroclor-1016-3       | (3) | 4.63       | 4.63      | 4.53      | 4.73 | 0.00       |
| Aroclor-1016-4       | (4) | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Aroclor-1016-5       | (5) | 5.01       | 5.01      | 4.91      | 5.11 | 0.00       |
| Aroclor-1260-1       | (1) | 6.11       | 6.12      | 6.02      | 6.22 | 0.01       |
| Aroclor-1260-2       | (2) | 6.37       | 6.38      | 6.28      | 6.48 | 0.01       |
| Aroclor-1260-3       | (3) | 6.73       | 6.73      | 6.63      | 6.83 | 0.00       |
| Aroclor-1260-4       | (4) | 6.95       | 6.95      | 6.85      | 7.05 | 0.00       |
| Aroclor-1260-5       | (5) | 7.26       | 7.26      | 7.16      | 7.36 | 0.00       |
| Tetrachloro-m-xylene |     | 3.44       | 3.44      | 3.34      | 3.54 | 0.00       |
| Decachlorobiphenyl   |     | 8.65       | 8.65      | 8.55      | 8.75 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/10/2024 Initial Calibration Date(s): 04/23/2024 04/23/2024

Continuing Calib Time: 02:05 Initial Calibration Time(s): 10:25 17:00

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 3.73       | 3.73      | 3.63      | 3.83 | 0.00       |
| Aroclor-1016-2       | (2) | 3.75       | 3.75      | 3.65      | 3.85 | 0.01       |
| Aroclor-1016-3       | (3) | 3.91       | 3.91      | 3.81      | 4.01 | 0.01       |
| Aroclor-1016-4       | (4) | 3.95       | 3.96      | 3.86      | 4.06 | 0.01       |
| Aroclor-1016-5       | (5) | 4.15       | 4.15      | 4.05      | 4.25 | 0.00       |
| Aroclor-1260-1       | (1) | 5.14       | 5.14      | 5.04      | 5.24 | 0.00       |
| Aroclor-1260-2       | (2) | 5.33       | 5.33      | 5.23      | 5.43 | 0.00       |
| Aroclor-1260-3       | (3) | 5.47       | 5.47      | 5.37      | 5.57 | 0.00       |
| Aroclor-1260-4       | (4) | 5.93       | 5.94      | 5.84      | 6.04 | 0.01       |
| Aroclor-1260-5       | (5) | 6.18       | 6.18      | 6.08      | 6.28 | 0.00       |
| Tetrachloro-m-xylene |     | 2.73       | 2.73      | 2.63      | 2.83 | 0.00       |
| Decachlorobiphenyl   |     | 7.50       | 7.50      | 7.40      | 7.60 | 0.00       |