

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/10/2024 Initial Calibration Date(s): 04/23/2024 04/23/2024

Continuing Calib Time: 04:32 Initial Calibration Time(s): 10:25 17:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.55	4.55	4.45	4.65	0.00
Aroclor-1016-2	(2)	4.57	4.57	4.47	4.67	0.00
Aroclor-1016-3	(3)	4.63	4.63	4.53	4.73	0.00
Aroclor-1016-4	(4)	4.72	4.72	4.62	4.82	0.00
Aroclor-1016-5	(5)	5.01	5.01	4.91	5.11	0.00
Aroclor-1260-1	(1)	6.12	6.12	6.02	6.22	0.01
Aroclor-1260-2	(2)	6.38	6.38	6.28	6.48	0.00
Aroclor-1260-3	(3)	6.73	6.73	6.63	6.83	0.00
Aroclor-1260-4	(4)	6.95	6.95	6.85	7.05	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.44	3.44	3.34	3.54	0.00
Decachlorobiphenyl		8.65	8.65	8.55	8.75	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/10/2024 Initial Calibration Date(s): 04/23/2024 04/23/2024

Continuing Calib Time: 04:32 Initial Calibration Time(s): 10:25 17:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.73	3.73	3.63	3.83	0.00
Aroclor-1016-2	(2)	3.75	3.75	3.65	3.85	0.01
Aroclor-1016-3	(3)	3.91	3.91	3.81	4.01	0.01
Aroclor-1016-4	(4)	3.95	3.96	3.86	4.06	0.01
Aroclor-1016-5	(5)	4.15	4.15	4.05	4.25	0.00
Aroclor-1260-1	(1)	5.14	5.14	5.04	5.24	0.00
Aroclor-1260-2	(2)	5.33	5.33	5.23	5.43	0.00
Aroclor-1260-3	(3)	5.47	5.47	5.37	5.57	0.00
Aroclor-1260-4	(4)	5.93	5.94	5.84	6.04	0.01
Aroclor-1260-5	(5)	6.18	6.18	6.08	6.28	0.00
Tetrachloro-m-xylene		2.73	2.73	2.63	2.83	0.00
Decachlorobiphenyl		7.50	7.50	7.40	7.60	0.01