

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/09/2024 Initial Calibration Date(s): 04/24/2024 04/24/2024

Continuing Calib Time: 13:21 Initial Calibration Time(s): 10:15 16:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-4	(4)	5.06	5.07	4.97	5.17	0.01
Aroclor-1016-5	(5)	5.37	5.38	5.28	5.48	0.01
Aroclor-1260-1	(1)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-2	(2)	6.87	6.88	6.78	6.98	0.02
Aroclor-1260-3	(3)	7.25	7.27	7.17	7.37	0.02
Aroclor-1260-4	(4)	7.50	7.51	7.41	7.61	0.02
Aroclor-1260-5	(5)	7.83	7.85	7.75	7.95	0.02
Tetrachloro-m-xylene		3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl		9.54	9.57	9.47	9.67	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/09/2024 Initial Calibration Date(s): 04/24/2024 04/24/2024

Continuing Calib Time: 13:21 Initial Calibration Time(s): 10:15 16:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.00	4.01	3.91	4.11	0.01
Aroclor-1016-2	(2)	4.02	4.03	3.93	4.13	0.01
Aroclor-1016-3	(3)	4.20	4.20	4.10	4.30	0.00
Aroclor-1016-4	(4)	4.25	4.25	4.15	4.35	0.00
Aroclor-1016-5	(5)	4.47	4.47	4.37	4.57	0.00
Aroclor-1260-1	(1)	5.55	5.56	5.46	5.66	0.01
Aroclor-1260-2	(2)	5.76	5.76	5.66	5.86	0.00
Aroclor-1260-3	(3)	5.91	5.92	5.82	6.02	0.01
Aroclor-1260-4	(4)	6.42	6.42	6.32	6.52	0.00
Aroclor-1260-5	(5)	6.68	6.69	6.59	6.79	0.01
Tetrachloro-m-xylene		2.89	2.90	2.80	3.00	0.01
Decachlorobiphenyl		8.13	8.13	8.03	8.23	0.00