

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL01 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103772.D Time Analyzed: 13:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.618	5.518	5.718	479.650	500.000	-4.1
Aroclor-1016-2	5.641	5.541	5.741	483.690	500.000	-3.3
Aroclor-1016-3	5.703	5.603	5.803	482.500	500.000	-3.5
Aroclor-1016-4	5.802	5.701	5.901	488.730	500.000	-2.3
Aroclor-1016-5	6.096	5.995	6.195	475.290	500.000	-4.9
Aroclor-1260-1	7.219	7.118	7.318	493.220	500.000	-1.4
Aroclor-1260-2	7.474	7.373	7.573	508.560	500.000	1.7
Aroclor-1260-3	7.834	7.732	7.932	524.480	500.000	4.9
Aroclor-1260-4	8.058	7.957	8.157	506.500	500.000	1.3
Aroclor-1260-5	8.379	8.278	8.478	520.820	500.000	4.2
Decachlorobiphenyl	10.205	10.102	10.302	55.330	50.000	10.7
Tetrachloro-m-xylene	4.449	4.351	4.551	50.160	50.000	0.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL01 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103772.D Time Analyzed: 13:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.660	4.571	4.771	491.900	500.000	-1.6
Aroclor-1016-2	4.679	4.589	4.789	480.900	500.000	-3.8
Aroclor-1016-3	4.853	4.763	4.963	467.560	500.000	-6.5
Aroclor-1016-4	4.895	4.805	5.005	462.230	500.000	-7.6
Aroclor-1016-5	5.106	5.016	5.216	470.580	500.000	-5.9
Aroclor-1260-1	6.129	6.041	6.241	453.280	500.000	-9.3
Aroclor-1260-2	6.314	6.226	6.426	459.610	500.000	-8.1
Aroclor-1260-3	6.467	6.379	6.579	452.930	500.000	-9.4
Aroclor-1260-4	6.935	6.848	7.048	467.490	500.000	-6.5
Aroclor-1260-5	7.174	7.087	7.287	479.660	500.000	-4.1
Decachlorobiphenyl	8.539	8.456	8.656	47.480	50.000	-5.0
Tetrachloro-m-xylene	3.584	3.491	3.691	51.680	50.000	3.4