

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103783.D Time Analyzed: 16:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.619	5.518	5.718	491.680	500.000	-1.7
Aroclor-1016-2	5.642	5.541	5.741	486.220	500.000	-2.8
Aroclor-1016-3	5.703	5.603	5.803	489.090	500.000	-2.2
Aroclor-1016-4	5.803	5.701	5.901	491.740	500.000	-1.7
Aroclor-1016-5	6.097	5.995	6.195	488.480	500.000	-2.3
Aroclor-1260-1	7.220	7.118	7.318	506.430	500.000	1.3
Aroclor-1260-2	7.476	7.373	7.573	518.520	500.000	3.7
Aroclor-1260-3	7.835	7.732	7.932	532.200	500.000	6.4
Aroclor-1260-4	8.060	7.957	8.157	515.650	500.000	3.1
Aroclor-1260-5	8.381	8.278	8.478	530.860	500.000	6.2
Decachlorobiphenyl	10.209	10.102	10.302	55.950	50.000	11.9
Tetrachloro-m-xylene	4.450	4.351	4.551	51.430	50.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103783.D Time Analyzed: 16:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.661	4.571	4.771	478.900	500.000	-4.2
Aroclor-1016-2	4.680	4.589	4.789	486.730	500.000	-2.7
Aroclor-1016-3	4.854	4.763	4.963	472.310	500.000	-5.5
Aroclor-1016-4	4.895	4.805	5.005	467.440	500.000	-6.5
Aroclor-1016-5	5.106	5.016	5.216	479.840	500.000	-4.0
Aroclor-1260-1	6.129	6.041	6.241	458.700	500.000	-8.3
Aroclor-1260-2	6.315	6.226	6.426	463.360	500.000	-7.3
Aroclor-1260-3	6.468	6.379	6.579	456.290	500.000	-8.7
Aroclor-1260-4	6.936	6.848	7.048	472.080	500.000	-5.6
Aroclor-1260-5	7.175	7.087	7.287	481.760	500.000	-3.6
Decachlorobiphenyl	8.541	8.456	8.656	48.320	50.000	-3.4
Tetrachloro-m-xylene	3.584	3.491	3.691	52.550	50.000	5.1